



The Safety Company

MSA Corporate Center • 1000 Cranberry Woods Drive • Cranberry Township, PA 16066

www.msasafety.com

Telephone: (800) MSA-2222

ALTAIR CERTIFICATE OF CALIBRATION

Serial Number: 08604150

Part Number: 10071361

Sales Order Number:

SAS AF INGENIERIE
24 rue Louis Blanc
F-75010 Paris



Factory Calibration Date: 08/19/16

Set Points

	H2S 0-100.0 PPM					
↓ (Low)	5.0 PPM					
↑ (High)	10.0 PPM					
STEL						
TWA						
Calibrated Value	H2S 34 PPM					
Cylinder Lot #	993965					

All applicable inspections, testing, and calibrations were performed using NIST traceable equipment, where available, in accordance with MSA's ISO 9001 Certified Quality System. Each material, component, and/or instrument must be installed, operated and maintained in strict accordance with its labels, cautions, warnings, instructions, and within the limitations stated in the supplied instruction manual. Routine calibration checks, equipment inspections, and applicable preventative maintenance measures must be performed to verify that the materials, components, and/or instruments are operating properly. Failure to perform these tasks on a routine basis, or suggested intervals, with specified equipment or methods, may result in inaccurate readings.

Process Certified By:

Calibrated By: C.Simpson

JIM HOFFMAN
QUALITY ENGINEER

LOCATION: 1000 Cranberry Woods Drive • Cranberry Township, PA 16066-5296